



UNITED STATES PATENT AND TRADEMARK OFFICE

m7

UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/088,843	07/12/2002	Philip Braithwaite	2245/109	2181
2101	7590	07/20/2004	EXAMINER	
BROMBERG & SUNSTEIN LLP 125 SUMMER STREET BOSTON, MA 02110-1618			DAWSON, GLENN K	
			ART UNIT	PAPER NUMBER
			3731	
DATE MAILED: 07/20/2004				

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary

Application No.

10/088,843

Applicant(s)

BRAITHWAITE

Examiner

Glenn K Dawson

Art Unit

3731

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 26 April 2004.
- 2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-20 is/are pending in the application.
- 4a) Of the above claim(s) _____ is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-20 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. § 119

- 12) ☒ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☒ All b) ☐ Some * c) ☐ None of:
1. ☒ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

* See the attached detailed Office action for a list of the certified copies not received.

Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO-1449 or PTO/SB/08)
Paper No(s)/Mail Date _____
- 4) ☐ Interview Summary (PTO-413)
Paper No(s)/Mail Date. _____
- 5) ☐ Notice of Informal Patent Application (PTO-152)
- 6) ☐ Other: _____

Specification

Applicant is reminded of the proper content of an abstract of the disclosure.

A patent abstract is a concise statement of the technical disclosure of the patent and should include that which is new in the art to which the invention pertains. If the patent is of a basic nature, the entire technical disclosure may be new in the art, and the abstract should be directed to the entire disclosure. If the patent is in the nature of an improvement in an old apparatus, process, product, or composition, the abstract should include the technical disclosure of the improvement. In certain patents, particularly those for compounds and compositions, wherein the process for making and/or the use thereof are not obvious, the abstract should set forth a process for making and/or use thereof. If the new technical disclosure involves modifications or alternatives, the abstract should mention by way of example the preferred modification or alternative.

The abstract should not refer to purported merits or speculative applications of the invention and should not compare the invention with the prior art.

Where applicable, the abstract should include the following:

- (1) if a machine or apparatus, its organization and operation;
- (2) if an article, its method of making;
- (3) if a chemical compound, its identity and use;
- (4) if a mixture, its ingredients;
- (5) if a process, the steps.

Extensive mechanical and design details of apparatus should not be given.

Applicant is reminded of the proper language and format for an abstract of the disclosure.

The abstract should be in narrative form and generally limited to a single paragraph on a separate sheet within the range of 50 to 150 words. It is important that the abstract not exceed 150 words in length since the space provided for the abstract on the computer tape used by the printer is limited. The form and legal phraseology often used in patent claims, such as "means" and "said," should be avoided. The abstract should describe the disclosure sufficiently to assist readers in deciding whether there is a need for consulting the full patent text for details.

The language should be clear and concise and should not repeat information given in the title. It should avoid using phrases which can be implied, such as, "The disclosure concerns," "The disclosure defined by this invention," "The disclosure describes," etc.

The abstract of the disclosure is objected to because it contains legal phraseology (means). Correction is required. See MPEP § 608.01(b).

Claim Rejections - 35 USC § 103

The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

The factual inquiries set forth in *Graham v. John Deere Co.*, 383 U.S. 1, 148 USPQ 459 (1966), that are applied for establishing a background for determining obviousness under 35 U.S.C. 103(a) are summarized as follows:

1. Determining the scope and contents of the prior art.
2. Ascertaining the differences between the prior art and the claims at issue.
3. Resolving the level of ordinary skill in the pertinent art.
4. Considering objective evidence present in the application indicating obviousness or nonobviousness.

Claims 1-6,8,9 and 11-20 are rejected under 35 U.S.C. 103(a) as being unpatentable over Braithwaite-Wo/93 16748 in view of Kimball, et al.-5944660, Barnes-'163 and Savolainen-Des. 389570.

Braithwaite discloses the inhaler as claimed including a reservoir means 44,21121,128133,145 having a dry powdered medicament therein. A metering device 123 in the form of a spool is also disclosed. However, the moisture-resistant coating on the reservoir and inhaler as a whole is not disclosed. Barnes discloses that it was known to coat the interior portions of an inhaler reservoir, mouthpiece and valve with

Art Unit: 3731

Parylene. Savolainen discloses that it was known to provide a moisture-resistant cover for a powder inhaler. Kimball discloses that it was known to provide a Parylene coating on devices as a moisture-resistant barrier. It would have been obvious to have coated the entire device as well as the powder reservoir of Braithwaite with a Parylene coating, as this has been shown to be an effective barrier against moisture, and would prevent the powder from absorbing moisture in the reservoir even after the desiccant is exhausted, and moisture from condensing on portions of the powder passageway, which produce undesirable clumping of the powder and degrades the ability of the metering member to operate correctly, thus producing inaccurate delivery of medicament doses to the user.

Claim 7 is rejected under 35 U.S.C. 103(a) as being unpatentable over Braithwaite-Wo/93 16748 in view of Kimball, et al.-5944660, Barnes-'163 and Savolainen-Des. 389570, as applied to the claims above, and further in view of Kawata-3798054.

Braithwaite as modified above makes obvious the invention as claimed with the exception of the coating being a sugar.

Kawata discloses that it was known to provide sugar as a moisture resistant coating.

It would have been obvious to have used a sugar as a moisture-resistant coating as sugars were known to have this characteristic.

Claim 10 is rejected under 35 U.S.C. 103(a) as being unpatentable over Braithwaite-Wo/93 16748 in view of Kimball, et al.-5944660, Barnes-'163 and

Art Unit: 3731

Savolainen-Des. 389570, as applied to the claims above, and further in view of Taylor-5067491

Braithwaite as modified above makes obvious the invention as claimed with the exception of the thickness of the coating. Taylor discloses that it was known to provide a Parylene moisture-resistant coating in the claimed thickness.

It would have been obvious to have made the coating within the claimed thickness, as the prior art has shown this thickness to be effective at providing a barrier against moisture.

Response to Arguments

Applicant's arguments with respect to claims 1-20 have been considered but are moot in view of the new ground(s) of rejection.

Conclusion

Applicant's amendment necessitated the new ground(s) of rejection presented in this Office action. Accordingly, **THIS ACTION IS MADE FINAL**. See MPEP § 706.07(a). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any

Art Unit: 3731

extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the date of this final action.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Glenn K Dawson whose telephone number is 703-308-4304. The examiner can normally be reached on M-Th 7:30-5:00.

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, McDermott or Shaver can be reached on 703-308-0858. The fax phone number for the organization where this application or proceeding is assigned is 703-872-9306.

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).



Glenn K Dawson
Primary Examiner
Art Unit 3731

Gkd
19 July 2004